

GOVERNMENT OF WEST BENGAL
Office of the Principal
NRS Medical College, Kolkata-14

E-mail: principalnrsmc@gmail.com

phone no : 2286-0033(direct)

Memo No: NMC/Store/ 2652

Date: 28/05/2019


Quotation Notice

Sealed quotations are invited from bonafide firms with superscription "Quotation for Ph Meter" for IVF centre under the Dept. of G & O of this Institution. The quotation should reach the office of the undersigned on or before 6.6.19 by 1 PM. The quotation should accompany current valid trade Licence, Income Tax Return, PAN No., GST Registration Certificate (not provisional)

Acceptance of the lowest rate is not obligatory for the undersigned who reserves the right to accept or reject any offer without assigning any reason whatsoever. All items should be good quality.

Rate must be submitted after calculating GST or other Taxes, if any according to the following format

| SI No. | Name Of Articles | RATE/UNIT |
|--------|------------------|-----------|
| 1 | Ph Meter | Per pc |



Principal

NRS Medical College, Kolkata-14

Memo No: NMC/Store / 2652 / 1 (8)

Dated: 28/05/2019

1. The Dy. Director of Health Services(E&S),Govt of West Bengal, CMS, Kolkata
2. The Joint Secretary, IT Cell, Department of Health and Family Welfare. Govt. Of West Bengal, Swasthya Bhawan, Kol-, with the request for publication of the Quotation Notice and Terms & Condition in the Govt Website.
- 3.-5. The Principal, Medical College/R.G.Kar Medical College/National Medical College, Kolkata
6. The Medical Superintendent-cum-Vice Principal, NRSMCH, Kolkata
7. The Director ,IPGME&R, Kolkata
8. Notice Board



Principal

NRS Medical College, Kolkata-14

TECHNICAL SPECIFICATION OF pH METER 01 No.

1. **General Features:**


1. pH Meter readings should have excellent calibration response and repeatability.
2. pH measuring range: 0.00- 14.0 Ph.
3. pH resolution range: 0.01-0.05.
4. pH accuracy: ± 0.05 .
5. Continuous ph monitoring (reading and recording every 30 minutes).
6. Track up to 5-10 trackpods. Real time ph monitoring and east data access.

2. Calibration:


1. Calibration: Digital automatic calibration and automatic off function.

3. Increased data storage capacity.

1. Data Export: Data export to PC with dedicated software.


22/2/19

Professor & Head
Dr. Chandana Das
Professor, Management & Obts
Dept. of Management & Obts
Dapur


21/2/19